


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2169, Rack A, Plate 1, Well C08, 501549
 [9969ADAB124F9B9D] Jun 05, 2026 14:16:25 Pressure = 64 bar (64 to 66)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.428 | F 0.51 | 0.9% | 17024 | |
| 2 | 0.508 | F 0.60 | 0.7% | 13782 | |
| 3 | 0.705 | F 0.84 | 17.3% | 331582 | Acetylated F peak |
| 4 | 0.832 | F 0.99 | 63.3% | 1215372 | Consistent with F |
| 5 | 1.000 | F 1.19 | 1.7% | 32867 | |
| 6 | 1.207 | A 1.02 | 15.9% | 304918 | A peak |
| 7 | 1.377 | S 0.88 | 0.2% | 4421 | |
| Total Area: | | | | 1919966 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range