


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2025, Rack A, Plate 1, Well F04, 171637
[9936C138B980A4E8] Dec 23, 2025 13:14:21 Pressure = 75 bar (75 to 77)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.402 | F 0.46 | 1.1% | 10397 | |
| 2 | 0.492 | F 0.57 | 0.8% | 7354 | |
| 3 | 0.705 | F 0.81 | 9.5% | 89994 | Acetylated F peak |
| 4 | 0.860 | F 0.99 | 54.6% | 516787 | Consistent with F |
| 5 | 0.995 | F 1.15 | 1.2% | 11612 | |
| 6 | 1.223 | A 1.01 | 32.8% | 310576 | A peak - REVIEW |
| Total Area: | | | | 946720 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range