


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2073, Rack A, Plate 1, Well A09, 172260
 [C927240E6304AEE8] Feb 13, 2026 12:47:57 Pressure = 53 bar (52 to 54)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.47	1.4%	15398	
2	0.517	F 0.59	0.5%	5228	
3	0.730	F 0.84	17.3%	187098	Acetylated F peak
4	0.882	F 1.01	63.6%	688956	Consistent with F
5	1.257	A 1.01	17.2%	186721	A peak
Total Area:				1083401	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range