


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2161, Rack A, Plate 1, Well A04, 501589  
[9D68E622D1D06C6A] May 29, 2026 11:14:12 Pressure = 62 bar (61 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.49	1.0%	5182	
2	0.505	F 0.60	0.5%	2586	
3	0.722	F 0.85	10.6%	57276	Acetylated F peak
4	0.843	F 1.00	61.7%	333723	Consistent with F
5	1.212	A 1.02	26.0%	140553	A peak - REVIEW
6	1.380	S 0.89	0.4%	1943	
Total Area:				541263	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range