


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2124, Rack A, Plate 1, Well G07, 502342  
[9D275CF65249FB95] Apr 16, 2026 12:52:11 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.48	1.0%	10111	
2	0.512	F 0.59	0.7%	7492	
3	0.715	F 0.82	12.6%	125796	Acetylated F peak
4	0.855	F 0.98	58.3%	584039	Consistent with F
5	1.240	A 0.99	27.3%	273666	A peak - REVIEW
Total Area:				1001104	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range