


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2191, Rack A, Plate 1, Well F08, 486035
[C839D42AE4D513BD] Jul 06, 2026 13:59:57 Pressure = 79 bar (79 to 81)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.50	1.3%	7709	
2	0.507	F 0.61	0.3%	2048	
3	0.672	F 0.81	3.7%	22087	Acetylated F peak
4	0.830	F 1.00	24.7%	145901	Consistent with F
5	0.975	F 1.17	3.8%	22211	
6	1.177	A 1.01	66.1%	390372	A peak - REVIEW
Total Area:				590328	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range