


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2029, Rack A, Plate 1, Well C03, 484135
 [8C6793FDC7F5F1BF] Dec 29, 2025 11:44:44 Pressure = 75 bar (74 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	0.6%	7948	
2	0.500	F 0.59	0.7%	8158	
3	0.695	F 0.82	13.2%	163379	Acetylated F peak
4	0.838	F 0.99	53.2%	656528	Consistent with F
5	1.000	F 1.18	1.6%	19725	
6	1.222	A 1.02	30.6%	377386	A peak - REVIEW
Total Area:				1233124	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range