


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2098, Rack A, Plate 1, Well C05, 498832  
 [C9367F0156B371B7] Mar 19, 2026 12:12:30 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.47	0.9%	9159	
2	0.518	F 0.58	0.7%	6653	
3	0.720	F 0.81	15.5%	158377	Acetylated F peak
4	0.858	F 0.97	67.4%	689277	Consistent with F
5	1.245	A 1.01	15.5%	158741	A peak
Total Area:				1022207	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range