


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2176, Rack A, Plate 1, Well F01, 493893
[883DA8142938EC40] Jun 12, 2026 13:54:11 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.423	F 0.50	0.9%	10049	
2	0.512	F 0.61	0.5%	5055	
3	0.703	F 0.83	11.7%	131068	Acetylated F peak
4	0.843	F 1.00	66.0%	739462	Consistent with F
5	1.183	A 1.00	20.9%	233958	A peak - REVIEW
Total Area:				1119592	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range