


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1957, Rack A, Plate 1, Well E01, 482279 FA
 [9D201F6AC63CDBBF] Oct 24, 2025 11:13:00 Pressure = 68 bar (68 to 70)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	1.0%	12101	
2	0.500	F 0.58	0.5%	6419	
3	0.715	F 0.83	17.2%	212621	
4	0.840	F 0.98	56.4%	697253	Consistent with F
5	1.005	F 1.17	1.7%	21485	
6	1.230	A 1.00	23.1%	286112	A peak - REVIEW
Total Area:				1235991	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range