


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2086, Rack A, Plate 1, Well C12, 492124
[9C233C717333713F] Mar 04, 2026 12:05:56 Pressure = 54 bar (54 to 56)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.47	1.5%	18087	
2	0.515	F 0.58	0.6%	7564	
3	0.730	F 0.83	17.6%	205786	Acetylated F peak
4	0.870	F 0.99	69.0%	809236	Consistent with F
5	1.018	F 1.16	1.4%	15894	
6	1.267	A 1.00	9.8%	115399	A peak
Total Area:				1171966	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range