


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1842, Rack A, Plate 1, Well E10, 456889 FA
 [8821D74154AA7117] Jun 25, 2025 13:14:10 Pressure = 61 bar (61 to 62)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.1%	14940	
2	0.497	F 0.58	1.4%	17821	
3	0.705	F 0.82	18.2%	237356	
4	0.832	F 0.97	60.3%	784810	Consistent with F
5	0.992	F 1.16	1.4%	18079	
6	1.223	A 1.01	17.5%	228059	A peak
Total Area:				1301065	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range