


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1970, Rack A, Plate 1, Well D06, 482300  
[D867F01C165A8ECA] Nov 07, 2025 11:38:32 Pressure = 67 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.46	1.0%	12572	
2	0.500	F 0.58	0.6%	7090	
3	0.720	F 0.84	14.3%	179749	Acetylated F peak
4	0.850	F 0.99	65.2%	819009	Consistent with F
5	1.000	F 1.16	0.8%	10177	
6	1.235	A 1.00	18.1%	227269	A peak
Total Area:				1255866	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range