


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2127, Rack A, Plate 1, Well D01, 502979
[8C364DCAC8E3F31D] Apr 21, 2026 10:59:29 Pressure = 62 bar (62 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.59	2.6%	30269	
2	0.720	F 0.83	9.9%	113227	Acetylated F peak
3	0.855	F 0.99	76.0%	868585	Consistent with F
4	1.248	A 1.01	11.4%	130243	A peak
Total Area:				1142324	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range