


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2188, Rack A, Plate 1, Well B05, 493029
[996DD48655A06C6A] Jul 01, 2026 11:24:52 Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.62	1.1%	14058	
2	0.698	F 0.83	14.1%	183174	Acetylated F peak
3	0.817	F 0.97	74.2%	964927	Consistent with F
4	1.167	A 1.00	10.6%	138105	A peak
Total Area:				1300264	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range