


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2061, Rack A, Plate 1, Well B08, 172031
[8D62CB10A94AAEE0] Jan 30, 2026 10:39:59 Pressure = 90 bar (90 to 92)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	0.6%	4332	
2	0.503	F 0.60	0.5%	3507	
3	0.693	F 0.83	13.1%	90022	Acetylated F peak
4	0.830	F 0.99	53.9%	370668	Consistent with F
5	0.987	F 1.18	2.7%	18916	
6	1.197	A 1.01	29.2%	200677	A peak - REVIEW
Total Area:				688122	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range