


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well C08, 486299
[D829811FBED5199D] Jul 09, 2026 11:23:53 Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.433	F	0.52	1.0%	9037	
2	0.518	F	0.63	0.2%	1824	
3	0.692	F	0.84	13.7%	125761	Acetylated F peak
4	0.830	F	1.00	68.6%	631139	Consistent with F
5	1.155	A	1.00	16.6%	152509	A peak
Total Area: 920270						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range