


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1964, Rack A, Plate 1, Well C12, 482206  
[9D73E490A2DAACC8] Oct 31, 2025 11:50:33 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	1.0%	13413	
2	0.502	F 0.58	0.5%	6113	
3	0.720	F 0.83	17.6%	226705	Acetylated F peak
4	0.848	F 0.98	71.3%	916770	Consistent with F
5	1.235	A 1.00	9.5%	122785	A peak
				Total Area: 1285786	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range