


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2156, Rack A, Plate 1, Well H10, 501508  
[9D2DB7D2F57044C2] May 26, 2026 14:40:42 Pressure = 65 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.60	1.8%	22439	
2	0.703	F 0.81	13.0%	159915	Acetylated F peak
3	0.848	F 0.98	62.2%	768403	Consistent with F
4	1.220	A 1.01	23.0%	283894	A peak - REVIEW
Total Area:				1234651	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range