


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1889, Rack A, Plate 1, Well D02, 169738
[CC313C894A4ED33D] Aug 20, 2025 12:06:45 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.0%	14007	
2	0.495	F 0.58	0.7%	10058	
3	0.700	F 0.83	14.6%	198353	Acetylated F peak
4	0.835	F 0.98	71.7%	974257	Consistent with F
5	1.220	A 1.01	11.9%	161422	A peak
Total Area:				1358097	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range