


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2124, Rack A, Plate 1, Well E03, 502338  
[DC6659E30D3E0EE0] Apr 16, 2026 11:53:35 Pressure = 62 bar (62 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.433	F 0.49	1.3%	12904	
2	0.515	F 0.59	1.1%	11320	
3	0.720	F 0.82	17.0%	172424	Acetylated F peak
4	0.872	F 0.99	71.8%	728938	Consistent with F
5	1.235	A 1.00	8.8%	89553	A peak
Total Area:				1015139	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range