


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1998, Rack A, Plate 1, Well C02, 483306 FA
 [CC32AD32F91A066A] Nov 28, 2025 11:39:00 Pressure = 70 bar (70 to 72)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.47	1.6%	18030	
2	0.487	F 0.57	2.3%	25682	
3	0.770	F 0.90	13.8%	150993	Acetylated F peak
4	0.857	F 1.00	59.8%	655245	Consistent with F
5	1.033	F 1.20	1.6%	17057	
6	1.253	A 1.02	20.9%	229052	A peak - REVIEW
Total Area:				1096059	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range