


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2098, Rack A, Plate 1, Well C08, 172738
[89777F0156B1F397] Mar 19, 2026 12:18:34 Pressure = 60 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.48	1.1%	10881	
2	0.515	F 0.58	0.7%	6889	
3	0.722	F 0.81	16.7%	166994	Acetylated F peak
4	0.870	F 0.98	67.5%	673947	Consistent with F
5	1.243	A 1.01	14.0%	140267	A peak
Total Area:				998978	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range