


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1359, Rack A, Plate 1, Well E02, 28163
[88D8E156DE5A6AE8] Apr 03, 2024 11:14:05 Pressure = 60 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.485	F 0.55	11.7%	168246	
2	0.882	F 1.00	78.8%	1131394	Consistent with F
3	1.250	A 1.00	9.5%	136575	A peak
Total Area:				1436215	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range