


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1970, Rack A, Plate 1, Well C08, 170938
[8932E45C1CD0AEC0] Nov 07, 2025 11:18:19 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.388	F 0.45	1.5%	9779	
2	0.495	F 0.57	0.6%	4292	
3	0.713	F 0.83	11.6%	76971	Acetylated F peak
4	0.855	F 0.99	54.7%	362263	Consistent with F
5	1.008	F 1.17	1.6%	10403	
6	1.237	A 1.00	30.0%	198409	A peak - REVIEW
Total Area:				662117	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range