


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2077, Rack A, Plate 1, Well F07, 503946
[D9777387A00EA4C8] Feb 18, 2026 12:03:53 Pressure = 55 bar (54 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.47	2.5%	16757	
2	0.513	F 0.58	1.0%	6795	
3	0.730	F 0.82	20.4%	136457	Acetylated F peak
4	0.885	F 1.00	69.6%	465085	Consistent with F
5	1.262	A 1.00	6.5%	43238	A peak
Total Area:				668332	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range