


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2161, Rack A, Plate 1, Well G10, 173393
[CC79A727A60D119D] May 29, 2026 13:55:53

Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.49	1.4%	22654	
2	0.502	F 0.59	0.9%	14356	
3	0.725	F 0.85	13.0%	217009	Acetylated F peak
4	0.835	F 0.98	56.3%	936755	Consistent with F
5	1.205	A 1.00	28.3%	470757	A peak - REVIEW
6	1.380	A 1.15	0.2%	3562	
Total Area: 1665093					

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range