


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2156, Rack A, Plate 1, Well B07, 501672
[DC78A7C7A8073B35] May 26, 2026 12:03:03 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.517	F	0.61	2.1%	21165	
2	0.720	F	0.84	17.2%	170693	Acetylated F peak
3	0.847	F	0.99	68.8%	681214	Consistent with F 6
4	1.222	A	1.01	11.8%	116454	A peak
Total Area: 989526						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range