


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2083, Rack A, Plate 1, Well G02, 492141  
 [9D322DE7AD997BBF] Feb 27, 2026 12:50:04 Pressure = 53 bar (52 to 54)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.46	1.0%	14711	
2	0.513	F 0.57	1.0%	13870	
3	0.738	F 0.82	15.7%	226594	Acetylated F peak
4	0.885	F 0.98	72.3%	1043279	Consistent with F
5	1.275	A 1.01	10.0%	143912	A peak
Total Area:				1442366	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range