


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1824, Rack A, Plate 1, Well B07, 456681
 [9861C4B1705D8CC8] Jun 04, 2025 11:17:29 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.48	1.1%	14624	
2	0.497	F 0.59	1.1%	15025	
3	0.723	F 0.86	18.4%	251304	Acetylated F peak
4	0.837	F 1.00	72.0%	985239	Consistent with F
5	1.215	A 1.01	7.4%	101626	A peak
Total Area:				1367818	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range