


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2161, Rack A, Plate 1, Well F06, 173570
 [C87DA632F9D8C668] May 29, 2026 13:23:33 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.50	0.7%	7074	
2	0.512	F 0.60	0.4%	4303	
3	0.700	F 0.83	13.6%	137367	Acetylated F peak
4	0.837	F 0.99	72.6%	732699	Consistent with F
5	1.207	A 1.00	12.6%	127179	A peak
Total Area:				1008622	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range