


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2046, Rack A, Plate 1, Well A09, 488974  
[CC77CC6C3937D13D] Jan 16, 2026 10:18:13 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.49	1.4%	8888	
2	0.498	F 0.59	1.0%	6153	
3	0.707	F 0.84	17.0%	109282	Acetylated F peak
4	0.848	F 1.01	62.7%	403346	Consistent with F
5	0.978	F 1.17	1.5%	9897	
6	1.200	A 1.01	16.5%	106152	A peak
Total Area:				643718	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range