


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2176, Rack A, Plate 1, Well C01, 502242
[9928B8417E45B915] Jun 12, 2026 12:37:23 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.432	F 0.51	1.4%	9706	
2	0.510	F 0.61	1.0%	6802	
3	0.717	F 0.85	17.7%	123769	Acetylated F peak
4	0.837	F 1.00	70.2%	490046	Consistent with F
5	1.193	A 1.01	9.7%	67997	A peak
Total Area:				698320	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range