


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1875, Rack A, Plate 1, Well A12, 169344  
 [C97121796E7986C0] Jul 30, 2025 11:17:47 Pressure = 59 bar (59 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.497	F 0.59	2.6%	60047	
2	0.707	F 0.84	15.3%	350929	Acetylated F peak
3	0.827	F 0.99	60.5%	1384925	Consistent with F
4	0.990	F 1.18	1.0%	21740	
5	1.222	A 1.01	20.3%	463777	A peak - REVIEW
6	1.380	A 1.14	0.3%	6233	
Total Area:				2287651	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range