


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1875, Rack A, Plate 1, Well D07, 169425  
 [C87434796C7BAC40] Jul 30, 2025 12:22:27 Pressure = 59 bar (59 to 60)

FA



| PEAK        | RT    | REL RT | % CONC | AREA   | COMMENT           |
|-------------|-------|--------|--------|--------|-------------------|
| 1           | 0.392 | F 0.46 | 1.5%   | 13812  |                   |
| 2           | 0.488 | F 0.58 | 1.4%   | 12452  |                   |
| 3           | 0.718 | F 0.85 | 10.4%  | 93087  | Acetylated F peak |
| 4           | 0.837 | F 0.99 | 62.8%  | 559942 | Consistent with F |
| 5           | 1.228 | A 1.01 | 23.5%  | 209520 | A peak - REVIEW   |
| 6           | 1.378 | A 1.14 | 0.3%   | 2339   |                   |
| Total Area: |       |        |        | 891152 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range