


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1860, Rack A, Plate 1, Well B07, 169007
[DC71CD089EC0D1B5] Jul 09, 2025 10:33:19 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.3%	12504	
2	0.492	F 0.58	1.1%	10428	
3	0.703	F 0.83	14.2%	134562	Acetylated F peak
4	0.860	F 1.01	66.7%	630767	Consistent with F
5	1.225	A 1.01	16.6%	156724	A peak
Total Area:				944985	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range