


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2029, Rack A, Plate 1, Well A09, 171664  
[8C32C7FDC7757137] Dec 29, 2025 11:08:22 Pressure = 76 bar (75 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.9%	5547	
2	0.500	F 0.59	0.7%	4037	
3	0.700	F 0.83	10.2%	60574	Acetylated F peak
4	0.845	F 1.00	53.6%	319795	Consistent with F
5	1.000	F 1.18	1.9%	11417	
6	1.218	A 1.01	32.7%	194726	A peak - REVIEW
Total Area:				596096	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range