


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1970, Rack A, Plate 1, Well A01, 482141
[DD36A15C1EFA8642] Nov 07, 2025 10:13:38 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.388	F 0.45	1.9%	11367	
2	0.497	F 0.58	0.6%	3637	
3	0.712	F 0.83	9.6%	58463	Acetylated F peak
4	0.862	F 1.00	46.6%	283439	Consistent with F
5	1.003	F 1.17	1.5%	9018	
6	1.103	A 0.90	0.4%	2667	
7	1.238	A 1.01	39.4%	239305	A peak - REVIEW
Total Area:				607896	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range