


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2068, Rack A, Plate 1, Well B03, 498918  
[CC36613B392E0CEA] Feb 10, 2026 10:51:56 Pressure = 97 bar (97 to 99)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.50	0.6%	5472	
2	0.508	F 0.61	0.4%	3242	
3	0.695	F 0.84	12.6%	113784	Acetylated F peak
4	0.828	F 1.00	58.9%	532822	Consistent with F
5	0.975	F 1.17	2.4%	22136	
6	1.188	A 1.01	25.1%	227006	A peak - REVIEW
Total Area:				904462	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range