


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2003, Rack A, Plate 1, Well F08, 171345 FA  
 [9D62AFA42547F397] Dec 03, 2025 12:27:06 Pressure = 72 bar (72 to 74)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	1.1%	12919	
2	0.498	F 0.58	0.4%	4476	
3	0.723	F 0.84	13.3%	162407	
4	0.847	F 0.99	68.1%	830424	Acetylated F peak Consistent with F
5	1.008	F 1.17	0.4%	5451	
6	1.238	A 1.00	16.7%	203008	A peak
Total Area:				1218685	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range