


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1964, Rack A, Plate 1, Well D01, 483373
[D972E490A2D82C40] Oct 31, 2025 11:52:35 Pressure = 66 bar (66 to 68)

AFU



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.385	F 0.45	2.1%	12214	
2	0.485	F 0.56	1.4%	7882	
3	0.705	F 0.82	5.9%	33439	Acetylated F peak
4	0.862	F 1.00	27.0%	153917	Consistent with F
5	1.022	F 1.18	4.4%	24890	
6	1.238	A 1.00	57.6%	327951	A peak - REVIEW
7	1.415	S 0.90	1.6%	9059	3
Total Area:				569352	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range