


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2077, Rack A, Plate 1, Well D03, 492107
[D96333D2F75151BF] Feb 18, 2026 11:05:16 Pressure = 56 bar (55 to 58)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|------|-------|--------|--------|--------|-------------------|
| 1 | 0.415 | F 0.47 | 1.5% | 20573 | |
| 2 | 0.513 | F 0.58 | 0.5% | 7344 | |
| 3 | 0.727 | F 0.83 | 17.9% | 240368 | Acetylated F peak |
| 4 | 0.865 | F 0.98 | 67.1% | 903665 | Consistent with F |
| 5 | 1.258 | A 1.00 | 12.9% | 173819 | A peak |

Total Area: 1345769

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range