


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2022, Rack A, Plate 1, Well C06, 171552  
[892694D320DF5B9D] Dec 19, 2025 11:06:33 Pressure = 75 bar (75 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	0.8%	6985	
2	0.500	F 0.58	0.5%	4409	
3	0.705	F 0.82	12.5%	107120	Acetylated F peak
4	0.847	F 0.99	69.3%	594821	Consistent with F
5	1.225	A 1.01	16.9%	144908	A peak
Total Area:				858243	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range