


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2133, Rack A, Plate 1, Well E05, 502930
[9D721AEC3A362E48] Apr 29, 2026 12:56:17 Pressure = 66 bar (65 to 67)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.515 | F 0.60 | 0.9% | 14341 | |
| 2 | 0.718 | F 0.84 | 15.2% | 239792 | Acetylated F peak |
| 3 | 0.860 | F 1.00 | 73.9% | 1169505 | Consistent with F |
| 4 | 1.225 | A 0.99 | 10.0% | 157910 | A peak |
| Total Area: | | | | 1581548 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range