


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2124, Rack A, Plate 1, Well H01, 173049
[8D3659A3073E8640] Apr 16, 2026 13:04:19 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.505	F 0.58	8.2%	40404	
2	0.735	F 0.84	20.3%	100431	
3	0.862	F 0.99	53.2%	263737	Consistent with F
4	1.013	F 1.16	2.7%	13294	
5	1.243	A 0.99	15.7%	77868	A peak
Total Area:				495734	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range