


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1905, Rack A, Plate 1, Well B10, 170085
 [C92040B787898442] Sep 10, 2025 11:39:14 Pressure = 61 bar (61 to 63)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.403 | F 0.49 | 0.9% | 9426 | |
| 2 | 0.495 | F 0.60 | 0.8% | 8580 | |
| 3 | 0.693 | F 0.84 | 13.1% | 133771 | Acetylated F peak |
| 4 | 0.823 | F 0.99 | 62.2% | 635379 | Consistent with F |
| 5 | 0.978 | F 1.18 | 1.0% | 9869 | |
| 6 | 1.207 | A 1.02 | 21.8% | 222430 | A peak - REVIEW |
| 7 | 1.367 | S 0.88 | 0.2% | 2384 | |
| Total Area: | | | | 1021839 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range