


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2022, Rack A, Plate 1, Well C01, 171543  
[DD6690D3207D5117] Dec 19, 2025 10:54:26 Pressure = 75 bar (75 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.495	F 0.58	3.9%	54835	
2	0.708	F 0.83	16.9%	238128	Acetylated F peak
3	0.847	F 1.00	64.4%	908660	Consistent with F
4	1.218	A 1.01	14.9%	209837	A peak
Total Area: 1411460					

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range