


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2061, Rack A, Plate 1, Well B11, 172035
 [9826DE45FC175197] Jan 30, 2026 10:46:02 Pressure = 90 bar (90 to 92)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.49	1.0%	11081	
2	0.500	F 0.60	0.6%	6383	
3	0.712	F 0.85	18.0%	208311	Acetylated F peak
4	0.827	F 0.99	65.3%	757664	Consistent with F
5	0.995	F 1.19	1.2%	13760	
6	1.203	A 1.02	14.0%	162269	A peak
Total Area:				1159468	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range