


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2188, Rack A, Plate 1, Well C07, 493253
 [DD7D80C32A7733B7] Jul 01, 2026 11:55:10 Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.437	F 0.52	0.7%	7837	
2	0.518	F 0.62	0.3%	3155	
3	0.693	F 0.83	13.8%	165274	Acetylated F peak
4	0.833	F 1.00	75.2%	902019	Consistent with F
5	1.162	A 1.01	10.0%	120468	A peak
Total Area:				1198753	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range