


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1988, Rack A, Plate 1, Well C03, 482227
 [C972B8E08D1086E2] Nov 21, 2025 11:30:38 Pressure = 70 bar (69 to 71)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.47	0.7%	11176	
2	0.502	F 0.59	0.6%	9495	
3	0.717	F 0.84	14.2%	226009	Acetylated F peak
4	0.862	F 1.01	69.2%	1101399	Consistent with F
5	1.232	A 1.01	15.3%	243184	A peak
Total Area:				1591263	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range